Prone Face Cushion, Adult
KYRA9080

1 Disconnect tubing and place cushion onto patient's face.



2 Place helmet over the cushion. Supporting the patient's face, head & neck, turn patient to prone position.



3 Lifting the helmet, slide mirror underneath and locate knobs in to helmet.



4 Check the patient's face is placed correctly with eyes clear of pressure.



5 Turn the adjustable knobs to achieve a neutral neckline.

For use with KYRA9071 Prone Head Support System

*Patent Pending

LOT

REF KYRA9080

REV B

CMP-2149-B

KYRA MEDICAL, LTD.
TWO ACRES,
HIGHER LANE, PRESTON
LANCASHIRE, PR4 6JH UK
+44 (0) 1772 370 544

KYRA MEDICAL, INC.
102 OTIS STREET
NORTHBOROUGH, MA
01532 USA
1-508-936-3550

EC REP
EU Authorised Representative:
Emergo Europe
Westervoortsedijk 60
6827 AT Arnhem
The Netherlands

EU Importer:
Kyra Medical Europe Limited.
69 Esker Woods Drive
Lucan Co Dublin, K78PX45
IRELAND

ISO 13485:2016 MD UK CA
LATEX NON STERILE CE

www.kyramedical.com

(01)5065024924003